

11-16-89

To - Helon Crawford

Subj - Evaluation of Proposed Technical  
Changes on Contract - NAS-9-18230

The contractor has proposed [redacted] very desirable changes in item 3, Isometric Force Measurement System - at no cost to the government. [redacted] which will [redacted] markedly improve performance and usability.

The changes consist of 1) replacing the velcro straps Item 3 Para. 1 with handles [redacted] for [redacted] for arm measurements, [redacted] for [redacted] gripping, and a stirrup for leg measurements and 2) replacing the planned [redacted] mechanical torque measurement system with an electronic measurement system with an electronic system with the same or superior performance.

The contractor has included construction details of the electronic system which will consist of an integral, battery operated unit with a remote wire connected digital readout.

No other aspects of the contract or apparatus will be affected.

I contacted SR & CA who requested

detailed

the information provided by the contractor on  
the ~~elect~~ proposed electronic unit fabrication

I have reviewed the proposed changes  
and the electronics unit details and I  
strongly recommend that these changes be  
made for the benefit of the government.

The contractor needs either an extension in  
delivery or a rapid response to this  
proposal to avoid a delay in delivery.

H. Truitt