

EXP. M092 COMMENT

NORMALLY THE SUBJECT IS ISOLATED THROUGH ELECTRO-SHOCK PROTECTION RETURN IMPEDENCE. LVMS LEG BANDS ARE CALIBRATED USING THIS GROUND RETURN IMPEDENCE. USE OF A WIRE TO STRUCTURE GROUND WILL RESULT IN A CHANGE IN CALIBRATION OF THE RIGHT LEG BAND. USE ELECTRODE DISC'S IF NECESSARY TO INSURE BETTER GROUND. RETURN THROUGH SUBJECT.

ABOVE INFO PER R. GOWEN (AFA)