

# USAF QUALITY CONTROL CORRECTIVE ACTION REQUEST

AF CORRECTIVE ACTION REQUEST NR

TO

FROM

1. DATE

2. SYSTEM OR AF INSPECTION STATION

3. THE FOLLOWING DISCREPANCIES REQUIRING CORRECTIVE ACTION HAVE BEEN FOUND BY AF QUALITY CONTROL PERSONNEL:

- (2) Test equipment not inspected prior to use. *correct*
- (3) Master gages standards of known accuracy not available. *correct*
- (5) Test equipment not inspected to required frequencies to maintain accuracies. *wrong no freq called*
- (101) Specifications or other technical data not available for control of test equipment. *correct*
- (102) Test equipment not identified. *correct*
- (103) Test equipment records or other evidence of control for test equipment not available.

4. IT IS REQUESTED THAT A REPORT OF CORRECTIVE ACTION TAKEN BE SUBMITTED TO THE AF QUALITY CONTROL REPRESENTATIVE WITHIN 15 DAYS OF THE ABOVE DATE. IN THE EVENT CORRECTIVE ACTION CANNOT BE FINALIZED BY THIS DATE, REQUEST DATE SAME WILL BE CONCLUDED. REVERSE SIDE MAY BE USED FOR YOUR REPLY. PLEASE REFER TO ABOVE AF C.A.R. NUMBER IF SEPARATE COMMUNICATION IS USED.

*To 7 Sullivan*

AIR FORCE QUALITY CONTROL