

CONCUR. Pls send to George Hoff
in procurement.
Carolyn
phys. copy this + send phys. to Hoff.
Joe Keenan

Joe it may not be proper for me to
comment on the proposal since one
company is suing Quinton for infringement
of one of my patents. (I am an innocent
& totally unreimbursed bystander in the matter).

If I were purchasing a tread mill
it would be either a Quinton or Collins
i.e. I have no reservations on the
quality & reputation of Quinton. If, as
I suspect, this unit is to be used
for clinical testing there seems little
justification for purchase of a research
unit $\dot{c}$ speeds to 15 MPH. The model
18-49C should be more than adequate, require
less space & installation costs. The
programmer & accessories are all
appropriate to an active clinical
testing program

St. Thornton