



*Don Rovers*

NATIONAL AERONAUTICS AND SPACE ADMINISTRATION  
MANNED SPACECRAFT CENTER  
HOUSTON, TEXAS 77058

REPLY TO  
ATTN OF:

CB

February 22, 1973

MEMORANDUM FOR RECORD

FROM: CB/W. Thornton

SUBJECT: Urine Volume Measurement by BMMD

The question of using mass to determine urine volume with the SL MMD's was recently raised. This has been advocated as a simple solution to the urine volume measurement problem since 1965. Even with poorly constrained bags with appreciable "slosh", accuracies of a few percent are possible and the original specifications for the BMMD required such accuracies.

If the urine is well constrained in rigid containers such as the present box and spring-loaded cover of the UVMS, high accuracies can be obtained as was recently demonstrated.

For the demonstration, three urine bags were filled with differing quantities of water. These were placed in the UVMS box with the VD cover in place and held to the front of the 1-g trainer BMMD by two SL "C" clamps. BMMD periods were recorded for each of the three bags plus box and clamps, box and clamps alone for tare, and a single point 900 gm calibration point made using the SMMD cal masses. The calculation of the water mass was made as shown on the enclosure, and the masses compared with weights obtained by a Metlar gravimetric balance.

<u>Liquid Sample</u>	<u>BMMD Mass/Gms</u>	<u>Gravimetric Mass/Gms</u>	<u>Error Gms</u>	<u>Error %</u>
1	1837.6	1836.0	+1.6	.08
2	764.5	763.0	+1.5	.20
3	2825.5	2822.0	+3.5	.12

These are the order of mass errors one should expect with reasonable care. There appears to be a small system error present. The gravimetric balance was not checked against the calibration masses, or this may represent a small amount of slosh.

The above errors must be added to those from variations in urine specific gravity. The average pooled daily specific gravities should be 1.005  $\pm$  .002 (individual voids may vary more than this and in extreme illness

one can get ranges of 1.020 to 1.000 though not for long). Using such a standard correction, one could reasonably expect errors of  $\pm 2\%$  from this source. If the sample bags were to be removed prior to mass measurement, then any variation in sample amount would produce an error. Temperature density variations are normally ignored. Urine bag variation would also have to be considered. I trust neither of these two items should be greater than a few grams.

The major concern at this time is not accuracy but crew time. If one crewman does all three measurements it should add about 30 minutes to the timeline.

*C. Oker*  
William Thornton, PE for M074/172

Addressees:  
GE/Don Robinson ✓  
DE4/R. McKinney  
CD2/J. Smotherman

Enclosure

CB/WThornton:cao:2/22/73:2321

Calculation of mass from BMMD period:

$T$  = period of three cycles: secs.

$M$  = mass: gms

$K_{1,2}$  = constants

$T \approx K M^{-2}$  and  $M \approx K_2 T^2$

Assuming: straight line form and taking zero point as BMMD period with tare wt. in place and the 900 gm cal mass as a second point on a straight line the slope is derived:

<u>Sample</u>	<u>T</u> <u>Sec</u>	<u>Av. T</u> <u>Sec</u>	<u>T<sup>2</sup></u>
Zero mass	2.93452 38 48 53 41	2.93446	8.611055
Zero mass +900 gms	3.00545 545 41	3.00543	9.032609

Slope:

$$\frac{\text{Mass}}{T^2 - T^2_{900}} = \frac{900}{9.03261 - 8.61141} = 2134.95 \text{ gm/sec}^2$$

$$\text{Calculating masses: } K_2 T_x^2 - K_2 T_o^2 = K_2 T_x^2 \times \frac{M}{T^2} = M$$

<u>Sample</u>	<u>T</u> <u>Sec</u>	<u>T Av.</u> <u>Sec</u>	<u>T<sup>2</sup></u> <u>Sec<sup>2</sup></u>	<u>BMMD Mass</u> <u>Gms</u>
1	2.99488 492 500 486	2.994915	8.96952	765.30
2	3.07786 777 780 768 770	3.077762	9.472618	1839.39
3	3.15216 199 214 216	3.152113	9.935816	2828.30



A small buoyancy correction must be made since gravimetric masses are increased by the amount they are buoyed up by air ( $\approx 1.3$  gm/1000 cc), and this effect is not present in the MMD. After taking into account the relative air displacements of steel cal mass and water samples, a correction of approximately  $-.1\%$  must be made. Making this correction, the weights are:

<u>Sample</u>	<u>Mass</u>
1	764.5
2	1837.6
3	2825.5