

Concluding remarks - With adequate and careful adherence to them procedures, the S/L SMMO should provide accuracies of considerably 2% or better for specimen measurements with occasional exceptions <sup>for</sup> ~~at~~ small samples. It will allow <sup>only</sup> a limited degree of exploration of the ultimate operational limitations of this method of mass measurement. At the moment measurement of temperature will probably be the limiting factor.